Title: Kodak logo Bitlit (Converted)-1
Creator: FreeHand 7.0
CreationDate: 6/10/97 11:14 AM

SERVICE BULLETIN

Health Imaging Products

Eastman Kodak Company, Health Imaging, Rochester, NY 14650

SERVICE BULLETIN NO. 286 **JUNE 1999**

3228 *Kodak X-Omat* 480 RA PROCESSOR

Modification No. 3 , Part No. 3H8729, Available

The purpose of Modification No. 3, is to improve reliability of the electronics and in particular the QUAD POWER SUPPLY, by reducing heat in the electrical box. This is accomplished by installing a BAFFLE between the DRYER and the ELECTRICAL BOX, which reduces heat from the DRYER entering the ELECTRICAL BOX. It also includes installing an improved QUAD POWER SUPPLY.

This is a Type 2, Selective Modification, but parts will be free until May 2000.

Assemblies No Longer Available

The 5600 BOARD ASSEMBLY, 748477 is no longer available as an assembly. Order the component parts:

Part No.	Description	Qty.	Notes
696156	5600 BD, PKD	1	Does contain U12, EPROM, P/N 990029 Does NOT contain ALARM, P/N 748517
963624	HOUSING	1	
914545	BRACKET-SWITCH, RIGHT	1	
914544	BRACKET-SWITCH, LEFT	1	
914599	SWITCH, REED ASY	2	

CIRCUIT BOARD DISPLAY ASSEMBLY, 981425 is no longer available as an assembly. Order the component parts:

Part No.	Description	Qty.
981032	350 BD, PKD	1
981016	LCD, PKD	1
914888	FAN	1

Deb Trubatch, Service Engineer- Rochester

ã Eastman Kodak Company

For More Information:

U.S. Distributors Contact:

Health Imaging
Customer Support Operations
1-800-336-4722

GCSS Personnel Contact:

Film Handling TAC
U.S.: 1-800-822-1414
Canada: 1-800-433-1414
International: 716-724-4675

Kodak and *X-Omat* are trademarks.

Pub No. 3E0805

sb3228_jun99_286.doc

HEALTH IMAGING

Printed in USA

EASTMAN KODAK COMPANY – ROCHESTER, NY 14650